



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/052,812	01/18/2002	Sergiu Silvan	A02P1003	3212

36802 7590 07/27/2004

PACESETTER, INC.
15900 VALLEY VIEW COURT
SYLMAR, CA 91392-9221

EXAMINER

EVANISKO, GEORGE ROBERT

ART UNIT	PAPER NUMBER
----------	--------------

3762

DATE MAILED: 07/27/2004

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

10/052,812

Applicant(s)

SILVIAN ET AL.

Examiner

George R Evanisko

Art Unit

3762

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 18 January 2002.
- 2a) ☐ This action is FINAL. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-25 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-15 and 19-25 is/are rejected.
- 7) ☒ Claim(s) 16-18 is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- ☒ Notice of References Cited (PTO-892)
- ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- ☒ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date 2.
- ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____.
- ☐ Notice of Informal Patent Application (PTO-152)
- ☐ Other: _____.

Art Unit: 3762

DETAILED ACTION

Claim Rejections - 35 USC § 112

The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Claims 3-9 are rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention.

In claim 3, “are periodically connected” makes the claim incomplete since it does not state what element is performing the periodic connection or controlling the connection.

In claim 5, “further comprises a second switching element” is vague. Is this the same element as the switching network in claim 1 or another switch? It is suggested to further limit the switching element (“wherein the switching element comprises...”).

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1, 20, and 23 are rejected under 35 U.S.C. 102(b) as being anticipated by Kroll et al (5620464). Kroll uses switch 76 to connect and disconnect the capacitor to primary and switch 83 to turn primary on and off to deliver the charge to delivery capacitor 32 for heart electrodes 40 and 41.

Art Unit: 3762

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

This application currently names joint inventors. In considering patentability of the claims under 35 U.S.C. 103(a), the examiner presumes that the subject matter of the various claims was commonly owned at the time any inventions covered therein were made absent any evidence to the contrary. Applicant is advised of the obligation under 37 CFR 1.56 to point out the inventor and invention dates of each claim that was not commonly owned at the time a later invention was made in order for the examiner to consider the applicability of 35 U.S.C. 103(c) and potential 35 U.S.C. 102(e), (f) or (g) prior art under 35 U.S.C. 103(a).

Claims 2-6, 10-15, 19, 21, 22, 24, and 25 are rejected under 35 U.S.C. 103(a) as being unpatentable over Kroll et al.

Kroll discloses the claimed invention having heart electrodes and first primary switch 83, except for the leads for the electrodes (claims 2, 21, and 22), the high frequency alternating signal to the first switch (claims 3, 4, 6, 12, 13, 15, and 24), the controller determining the state of charge of the delivery capacitor to enter the charging cycle or to stop charging the bypass capacitor (claims 11 and 24), and the bypass capacitor being ceramic capacitors and the delivery capacitor being electrolytic (claims 10 and 19). It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify the defibrillator system as taught by Kroll, with leads for the electrodes, the high frequency alternating signal to the first

Art Unit: 3762

switch, the controller determining the state of charge of the delivery capacitor to enter the charging cycle or to stop charging the bypass capacitor, and the bypass capacitor being ceramic capacitors and the delivery capacitor being electrolytic since it was known in the art that defibrillator systems use: leads for the electrodes so the defibrillator can be placed in a different location in the body but still provide the pulses directly to the heart; the high frequency alternating signal to the first switch to allow the primary to deliver charge to the secondary of the transformer and thereby deliver the energy to the delivery capacitor; the controller determining the state of charge of the delivery capacitor to enter the charging cycle or to stop charging the bypass capacitor to only use battery power when the capacitors are in need of charging; and the bypass capacitor being ceramic capacitors to provide capacitors that are free of outgassing, don't need reforming, and have a low resistance, and the delivery capacitor being electrolytic to provide a high energy density per unit volume and low leakage current capacitor.

Allowable Subject Matter

Claims 7-9 would be allowable if rewritten to overcome the rejection(s) under 35 U.S.C. 112, second paragraph, set forth in this Office action and to include all of the limitations of the base claim and any intervening claims.

Claims 16-18 are objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

Conclusion


The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.

Art Unit: 3762

Any inquiry concerning this communication or earlier communications from the examiner should be directed to George R Evanisko whose telephone number is 703 308-2612. The examiner can normally be reached on M-F 6:30-5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Angela Sykes can be reached on 703 308-5181. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).


George R Evanisko
Primary Examiner
Art Unit 3762

7/26/04

GRE
July 26, 2004